



Microscopic Form Measurement System UMAP Vision System



Mitutoyo


Extremely small styli down to ø15 µm make contact measurement of microscopic form a reality

The UMAP Vision System is an ultra-low-measuring-force probe that uses Mitutoyo's proprietary sensing technology.

With a lineup of styli ranging from a minimum diameter of 15 µm, this product responds to the needs of our customers for microscopic dimension and form measurement.




UMAP Vision System Features



Ultra low measuring force of a minimum 1 µN (UMAP103) makes it possible to measure even workpieces that are easily deformed. Up to three types of UMAP styli, each of a different diameter can be used in combination. The user can install, remove, and replace styli.



Non-contact measurement in vision mode and contact measurement using the UMAP mode can be done solely by one unit. Even for parts that are difficult to see, vision mode can be used to position workpiece, and then the UMAP mode can be used to perform aimed measurements.






Stylus with a minimum diameter of 15 µm makes contact measurements of microscopic areas possible.



Mitutoyo


TECHNOLOGY

Detection Principle




UMAP Stylus Lineup

Five types of stylus modules that differ in stylus tip radius and shaft length are available. Up to three types of styli can be used in combination, which makes it possible to select the optimum stylus to match the workpiece to be measured.




Application Examples

The following are examples of workpieces that can be measured with UMAP. UMAP provides a solution for microscopic dimension measurement and form evaluation in a wide variety of fields such as precise microscopic molds, micromachining pressed products, microscopic resin molded products, and EDM electrodes.



Micro connector molds and electrode measurement



Lens barrel measurement



1. As shown in the figure to the left, the stylus maintains a microvibrating

2. When the stylus comes into contact with the workpiece, the stylus

3. When the stylus moves away from the workpiece, the stylus vibration amplitude returns to the state that was present when the stylus was not in contact with the workpiece, which makes the stylus ready for use in

vibration is restricted by the workpiece, which causes the vibration

amplitude to decrease. A touch-trigger signal will be detected when the

state when not in contact with the workpiece.

amplitude decreases past a certain level.


performing the next measurement.

Measurement of microscopic gear tooth and tooth-form profile / tooth-alignment profile

SOFTWARE

OVPAK

QVPAK, the main software, supports both non-contact measurements using vision mode and contact measurements using UMAP mode.



Calculation function example $[\bullet / \bigcirc = \bigcirc]$


Switching between vision and UMAP modes

The user can click icons to switch between vision and UMAP modes. This makes it possible to use vision mode to perform workpiece positioning and then use UMAP mode to perform targeted measurements.



Outlier removal


Outliers caused by dirt, burrs, and chips are removed automatically. Also, it is possible to determine the optimum removal level for abnormal points while viewing the measured data.




OPTIONAL SOFTWARE FORMTRACEPAK-AP

Form Evaluation and Analysis Software

Verifies designed values and performs form analysis on the basis of the contour data obtained via the UMAP or vision mode








Measuring tool arrangement with a single click

The tool size, orientation, and threshold of vision measuring tools are automatically set with one click of the mouse in the vicinity of the measurement location.



QVGraphics


The elements measured with QVPAK and the measurement results can be displayed graphically. It is also possible to calculate distances and angles for configuring the coordinate system settings of displayed elements that the user specifies.



Drawing the geometrical deviation of a circle Graphical display of a measured element



Analysis on microscopic dimensions



Gear contour matching and over-pin diameter analysis

Mitutoyo

OPTIONS



QV Objectives

Objective		QV-SL0.5x*1	QV-HR1x	QV-SL1x	QV-HR2.5x	QV-SL2.5x	QV-HR5x	QV-5x	QV-HR10x*1	QV-10x*1	QV-25x*1
Code No.		02AKT199	02AKT250	02ALA150	02AKT300	02ALA170	02AWD010	02ALA420	02AKT650	02ALG010	02ALG020
Set of objectives that support PFF		-	-	-	02AKX895B	-	02AXA915B	02AKX900B	02AKX905B	-	02AKX910B
Working distance [mm]		30.5	40.6	52.5	40.6	60.0	20.0	33.5	20.0	30.5	13.0
Field of view [mm]* ² (H)×(V)	Turret 1x	12.54x9.4	6.27x4.7		2.49	(1.86	1.24	(0.93	0.62>	(0.47	0.25x0.18
	Turret 2x	6.27x4.7	3.13x2.35		1.24x0.93		0.62x0.47		0.31x0.23		0.12x0.09
	Turret 6x	2.09x1.56	1.04x0.78		0.41x0.31		0.20	(0.15	0.10>	(0.07	0.04x0.03
	Digital magnificationg 12x	1.04x0.78	0.52x0.39		0.20x0.15		0.10	(0.07	0.05>	(0.03	0.02x0.01

*1 When the QV-SL0.5x, QV-HR10x, QV-10x, or QV-25x objective is used, some limitations, such as the illumination being insufficient depending on the workpiece, may occur. *2 The values for field of view are not guaranteed values, but representative values.

Calibration

Calibration gauge (ring) Calibration gauge (ø1 ball)



Alignment adjusting equipment



LINE-UP



UMAP Vision System TYPE2 HYPER 302 Pro




UMAP Vision System TYPE2 ULTRA 404


Specifications

Model name			UMAP Vision System TYPE2 HYPER 302 Pro	UMAP Vision System TYPE2 ULTRA 404				
Model code			UVS2-H302P1L-E	UVS2-U404P1L-E				
Main unit								
Measuring range (X×Y×Z) [mm] Vision Vision+UMAP		Vision	300×200×200	400×400×200 Effective measuring range when contour light is used:360×400×20				
		Vision+UMAP	185×200×175	285×400×175				
Stage glass (W×D) [mm]			399×271	493×551				
External dimensions (D×W×H) (Including dedicated stand) [mm]			930×866×1612	1736×1172×1910				
Main unit mass (Including dedicated stand) [kg]		ted stand) [kg]	365	2185				
Imaging device			B&W CMOS digital camera					
Observation unit*1			Programmable power turret 1x, 2x, 6x, (12x)					
	Vertical coaxial light		White LED	White LED				
Illumination unit	Stage light		White LED	White LED				
Programma		e ring light	White LED	White LED				
Maximum stage loading [kg]*2			15 kg	40 kg				
Vision measuring accuracy [µm]* ³		EUX,MPE / EUY,MPE	(1) (0.8 + 2 L/1000)µm / ② (0.8 + 3L/1000)	(1) (0.35 + 1.3 L/1000)µm / (2) (0.35 + 1.8 L/1000)				
	Vision*4	Euz,mpe	(1,5 + 2 L/1000)µm / ② (1.5 + 3.5L/1000)	1 (1.5 + 2 L/1000)µm / 2 (1.5 + 3.5 L/1000)				
		EUXY,MPE	 (1.4 + 3 L/1000)µm / ② (1.4 + 4L/1000) 	① (0.5 + 2 L/1000)µm / ② (0.5 + 2.5 L/1000)				
	STREAM	E1X, E1Y	(1.5 + 3 L/1000)	-				
	(optional)	E2XY	(2.0 + 4 L/1000)	-				
	Optical magnification		2.5x objective (QV-HR2.5x or QV-SL2	.5x) and middle magnification tube lens				
	UMAP*5	E1X, E1Y	(1.7 + 3 L/1000)	(1.5 + 3 L/1000)				
Dava a stark i lite (f. 1991)	Dimensional measurement within the FOV		-	3σ≤0.2				
	UMAP*3 *6	UMAP101, 103, 107	σ≤0.1	σ≤0.08				
		UMAP110, 130	σ≤0.15	σ≤0.12				
Operating	Ambient tem	perature	① 20±2 ℃ / ② 19 to 24 ℃					
temperature range*7 Temperature variation		variation	0.5 °C/1H and 1 °C/24H					
Thermal compensation function			Automatic					
Air cupphy	Pressure		0.4 MPa (Air supply pressure to be in range 0.5 to 0.9 MPa.)					
Air supply	Required air flow rate*8		_	300 L/min (ANR) Operating air pressure				

*1 Programmable power turret 1x, 2x, 4x, 6x model is available by special order. Digital magnification function allows 8x and 12x in addition to 1x, 2x, 4x, 6x. 6 steps of magnification are available in total. *2 An excessively biased or concentrated load is excluded.

*3 L: arbitrary measuring length (unit: mm) *4 According to ISO 10360-7:2011 (JIS B 7440-7:2015)







*5 The assured accuracy of UMAP is specific to that of UMAP110 in the case of a measuring speed of 10 μ m/s.

*6 The accuracies are guaranteed for a measuring speed of 5 μm/s for the UMAP101 and a measuring speed of 10 μm/s for the UMAP103, 107, 110, and 130.
*7 The guaranteed accuracy depends on ambient temperature.

*8 HYPER UMAP Vision System 302 TYPE2 HYPER 302 Pro only uses air to move UMAP up and down.



Whatever your challenges are, Mitutoyo supports you from start to finish.

Mitutoyo is not only a manufacturer of top quality measuring products but one that also offers qualified support for the lifetime of the equipment, backed up by comprehensive services that ensure your staff can make the very best use of the investment.

Apart from the basics of calibration and repair, Mitutoyo offers product and metrology training, as well as IT support for the sophisticated software used in modern measuring technology. We can also design, build, test and deliver measuring solutions and even, if deemed cost-effective, take your critical measurement challenges in-house on a sub-contract basis.



Find additional product literature and our product catalogue

https://www.mitutoyo.co.jp/global.html

Our products are classified as regulated items under Japanese Foreign Exchange and Foreign Trade Law. Please consult us in advance if you wish to export our products to any other country. If the purchased product is exported, even though it is not a regulated item (Catch-All controls item), the customer service available for that product may be affected. If you have any questions, please consult your local Mitutoyo sales office.

Note: Product illustrations are without obligation. Product descriptions, in particular any and all technical specifications, are only binding when explicitly agreed upon.

MITUTOYO and MiCAT are either registered trademarks or trademarks of Mitutoyo Corp. in Japan and/or other countries/regions. Other product, company and brand names mentioned herein are for identification purposes only and may be the trademarks of their respective holders.

All product information contained in this brochure is current as of July 2023.



Mitutoyo Corporation

20-1, Sakado 1-Chome, Takatsu-ku, Kawasaki-shi, Kanagawa 213-8533, Japan T +81 (0) 44 813-8230 F +81 (0) 44 813-8231 https://www.mitutoyo.co.jp